

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082721\
 Data File : VX023981.D
 Acq On : 27 Aug 2021 10:05
 Operator : JC/MD
 Sample : VX0827MBL01
 Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0827MBL01

Quant Time: Aug 28 04:53:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	175067	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	292791	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	258559	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	115016	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	146509	59.962	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	119.920%#
35) Dibromofluoromethane	5.397	113	106402	54.906	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.820%
50) Toluene-d8	8.653	98	376945	51.106	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.220%
62) 4-Bromofluorobenzene	11.085	95	134681	49.602	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082721\
Data File : VX023981.D
Acq On : 27 Aug 2021 10:05
Operator : JC/MD
Sample : VX0827MBL01
Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0827MBL01

Quant Time: Aug 28 04:53:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

