

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082721\
 Data File : VX023984.D
 Acq On : 27 Aug 2021 12:00
 Operator : JC/MD
 Sample : M3547-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK-8-25-21

Quant Time: Aug 28 04:54:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	132355	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	225486	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	209818	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	84844	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	95292	51.586	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.180%
35) Dibromofluoromethane	5.391	113	72712	48.720	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.440%
50) Toluene-d8	8.653	98	271466	47.791	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.580%
62) 4-Bromofluorobenzene	11.085	95	94236	45.066	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.140%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	2594	3.141	ug/l	97
17) Carbon Disulfide	2.514	76	1406	0.464	ug/l	95
20) Methylene Chloride	2.794	84	2457	1.557	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082721\
Data File : VX023984.D
Acq On : 27 Aug 2021 12:00
Operator : JC/MD
Sample : M3547-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK-8-25-21

Quant Time: Aug 28 04:54:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

