

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082721\
 Data File : VX023986.D
 Acq On : 27 Aug 2021 12:47
 Operator : JC/MD
 Sample : M3535-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WVLY-SW-12-GW-082421

Quant Time: Aug 28 04:55:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	138521	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	238469	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	222342	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90060	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	103508	53.539	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.080%	
35) Dibromofluoromethane	5.397	113	80244	50.840	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.680%	
50) Toluene-d8	8.653	98	288782	48.071	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 96.140%	
62) 4-Bromofluorobenzene	11.085	95	99876	45.163	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 90.320%	
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.172	85	959	0.656	ug/l	91
7) Trichlorofluoromethane	1.892	101	330058	133.573	ug/l	98
9) 1,1,2-Trichlorotrifluo...	2.331	101	2467	1.774	ug/l	93
12) 1,1-Dichloroethene	2.324	96	2788	2.097	ug/l	92
16) Acetone	2.392	43	2440	2.823	ug/l #	77
27) cis-1,2-Dichloroethene	4.501	96	2592	1.534	ug/l	87
30) Chloroform	5.104	83	77746	27.038	ug/l	96
32) 1,1,1-Trichloroethane	5.403	97	914	0.371	ug/l #	46
44) Trichloroethene	7.135	130	47615	26.396	ug/l	99
47) Bromodichloromethane	7.830	83	2371	1.086	ug/l	98
64) Tetrachloroethene	9.281	164	6410	3.610	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082721\
Data File : VX023986.D
Acq On : 27 Aug 2021 12:47
Operator : JC/MD
Sample : M3535-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WVLY-SW-12-GW-082421

Quant Time: Aug 28 04:55:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

