

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082721\
 Data File : VX023988.D
 Acq On : 27 Aug 2021 13:35
 Operator : JC/MD
 Sample : M3561-21
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK-082621

Quant Time: Aug 28 04:55:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	139895	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	239725	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	221632	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	84228	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	98426	50.410	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.820%	
35) Dibromofluoromethane	5.397	113	75207	47.399	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 94.800%	
50) Toluene-d8	8.653	98	278405	46.101	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 92.200%	
62) 4-Bromofluorobenzene	11.085	95	93360	41.995	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 84.000%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	2563	2.936	ug/l #	86
20) Methylene Chloride	2.794	84	1475	0.885	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082721\
Data File : VX023988.D
Acq On : 27 Aug 2021 13:35
Operator : JC/MD
Sample : M3561-21
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK-082621

Quant Time: Aug 28 04:55:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

