

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082721\
 Data File : VX023991.D
 Acq On : 27 Aug 2021 14:45
 Operator : JC/MD
 Sample : M3561-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OUII-C

Quant Time: Aug 28 04:56:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	134851	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	230330	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	214137	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85667	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	100395	53.342	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 106.680%			
35) Dibromofluoromethane	5.397	113	77614	50.911	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.820%			
50) Toluene-d8	8.653	98	283966	48.940	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 97.880%			
62) 4-Bromofluorobenzene	11.085	95	97179	45.496	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 91.000%			
Target Compounds						
16) Acetone	2.392	43	2516	2.990	ug/l	99
44) Trichloroethene	7.129	130	1156	0.663	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082721\
Data File : VX023991.D
Acq On : 27 Aug 2021 14:45
Operator : JC/MD
Sample : M3561-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OUII-C

Quant Time: Aug 28 04:56:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

