

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082721\
 Data File : VX023992.D
 Acq On : 27 Aug 2021 15:09
 Operator : JC/MD
 Sample : M3561-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OUII-E

Quant Time: Aug 28 04:57:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	130071	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	221281	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	203916	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	77413	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	96406	53.105	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.220%
35) Dibromofluoromethane	5.397	113	73803	50.391	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.780%
50) Toluene-d8	8.653	98	270570	48.538	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.080%
62) 4-Bromofluorobenzene	11.085	95	89459	43.594	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.180%
Target Compounds						
16) Acetone	2.392	43	2421	2.983	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082721\
Data File : VX023992.D
Acq On : 27 Aug 2021 15:09
Operator : JC/MD
Sample : M3561-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OUII-E

Quant Time: Aug 28 04:57:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

