

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082721\
 Data File : VX023993.D
 Acq On : 27 Aug 2021 15:33
 Operator : JC/MD
 Sample : M3561-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VE1-4

Quant Time: Aug 28 04:57:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	135561	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	234998	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	224619	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96709	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	101895	53.856	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.720%	
35) Dibromofluoromethane	5.397	113	78281	50.329	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.660%	
50) Toluene-d8	8.653	98	289283	48.866	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 97.740%	
62) 4-Bromofluorobenzene	11.085	95	103179	47.346	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 94.700%	
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	3094	3.657	ug/l	# 74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082721\
Data File : VX023993.D
Acq On : 27 Aug 2021 15:33
Operator : JC/MD
Sample : M3561-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VE1-4

Quant Time: Aug 28 04:57:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

