

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082721\
 Data File : VX023994.D
 Acq On : 27 Aug 2021 15:56
 Operator : JC/MD
 Sample : M3561-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUPLICATE-082521

Quant Time: Aug 28 04:57:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	148038	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	254661	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	238861	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	102875	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	107762	52.156	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	104.320%	
35) Dibromofluoromethane	5.397	113	81795	48.528	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	97.060%	
50) Toluene-d8	8.653	98	298353	46.507	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	93.020%	
62) 4-Bromofluorobenzene	11.085	95	106538	45.112	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	90.220%	
Target Compounds						
16) Acetone	2.386	43	2846	3.081	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082721\
Data File : VX023994.D
Acq On : 27 Aug 2021 15:56
Operator : JC/MD
Sample : M3561-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUPLICATE-082521

Quant Time: Aug 28 04:57:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

