

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082721\
 Data File : VX023995.D
 Acq On : 27 Aug 2021 16:20
 Operator : JC/MD
 Sample : M3561-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VE1-2

Quant Time: Aug 28 04:57:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	139255	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	238248	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	223076	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99793	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	101945	52.453	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.900%
35) Dibromofluoromethane	5.398	113	79981	50.720	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.440%
50) Toluene-d8	8.653	98	292912	48.804	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.600%
62) 4-Bromofluorobenzene	11.086	95	106460	48.185	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.360%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	4184	4.815	ug/l	90
17) Carbon Disulfide	2.520	76	2794	0.877	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082721\
Data File : VX023995.D
Acq On : 27 Aug 2021 16:20
Operator : JC/MD
Sample : M3561-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VE1-2

Quant Time: Aug 28 04:57:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

