

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082721\
 Data File : VX023996.D
 Acq On : 27 Aug 2021 16:43
 Operator : JC/MD
 Sample : M3561-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VE2-1

Quant Time: Aug 28 04:58:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	138116	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	236700	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	221385	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92540	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	101982	52.904	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.800%
35) Dibromofluoromethane	5.397	113	79036	50.449	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.900%
50) Toluene-d8	8.653	98	291165	48.830	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.660%
62) 4-Bromofluorobenzene	11.085	95	101169	46.089	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.180%
Target Compounds						
16) Acetone	2.385	43	4345	5.041	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082721\
Data File : VX023996.D
Acq On : 27 Aug 2021 16:43
Operator : JC/MD
Sample : M3561-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VE2-1

Quant Time: Aug 28 04:58:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

