

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082721\
 Data File : VX023998.D
 Acq On : 27 Aug 2021 17:31
 Operator : JC/MD
 Sample : M3561-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VE4-7

Quant Time: Aug 28 04:58:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	156355	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	267621	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	254020	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	120084	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.971	65	111864	51.262	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 102.520%		
35) Dibromofluoromethane	5.397	113	85191	48.095	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 96.180%		
50) Toluene-d8	8.653	98	322498	47.836	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 95.680%		
62) 4-Bromofluorobenzene	11.085	95	124552	50.186	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 100.380%		
Target Compounds						
						Qvalue
16) Acetone	2.392	43	9956	10.204	ug/l	88
17) Carbon Disulfide	2.520	76	2415	0.675	ug/l #	94
25) 2-Butanone	4.581	43	4072	2.755	ug/l	96
52) Toluene	8.720	92	2112	0.445	ug/l	100
68) m/p-Xylenes	10.305	106	5350	1.517	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082721\
Data File : VX023998.D
Acq On : 27 Aug 2021 17:31
Operator : JC/MD
Sample : M3561-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VE4-7

Quant Time: Aug 28 04:58:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

