

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082721\
 Data File : VX023980.D
 Acq On : 27 Aug 2021 09:42
 Operator : JC/MD
 Sample : VX0827WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0827WBL01

Quant Time: Aug 28 04:52:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	179967	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	291970	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	268212	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	125089	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	142916	56.899	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.800%
35) Dibromofluoromethane	5.391	113	105760	54.728	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.460%
50) Toluene-d8	8.653	98	378926	51.519	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.040%
62) 4-Bromofluorobenzene	11.085	95	139846	51.649	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.300%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082721\
Data File : VX023980.D
Acq On : 27 Aug 2021 09:42
Operator : JC/MD
Sample : VX0827WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0827WBL01

Quant Time: Aug 28 04:52:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 733
Delta R.T. -0.006 min
Lab File: VX023980.D
Acq: 27 Aug 2021 09:42




Instrument :
MSVOA_X
ClientSampleId :
VX0827WBL01


Tgt Ion:168 Resp: 179967
Ion Ratio Lower Upper
168 100
99 58.9 45.8 68.6



#33
1,2-Dichloroethane-d4
Concen: 56.899 ug/l
RT: 5.958 min Scan# 799
Delta R.T. -0.006 min
Lab File: VX023980.D
Acq: 27 Aug 2021 09:42

Tgt Ion: 65 Resp: 142916
Ion Ratio Lower Upper
65 100
67 50.3 0.0 105.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 931
Delta R.T. -0.006 min
Lab File: VX023980.D
Acq: 27 Aug 2021 09:42




Instrument :
MSVOA_X
ClientSampleId :
VX0827WBL01


Tgt Ion:114 Resp: 291970
Ion Ratio Lower Upper
114 100
63 21.9 0.0 40.8
88 16.2 0.0 32.8



#35
Dibromofluoromethane
Concen: 54.728 ug/l
RT: 5.391 min Scan# 706
Delta R.T. -0.006 min
Lab File: VX023980.D
Acq: 27 Aug 2021 09:42

Tgt Ion:113 Resp: 105760
Ion Ratio Lower Upper
113 100
111 103.1 81.9 122.9
192 18.7 16.7 25.1









#50
Toluene-d8
Concen: 51.519 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX023980.D
Acq: 27 Aug 2021 09:42




Instrument :
MSVOA_X
ClientSampleId :
VX0827WBL01


Tgt Ion: 98 Resp: 378926
Ion Ratio Lower Upper
98 100
100 64.0 51.9 77.9



#62
4-Bromofluorobenzene
Concen: 51.649 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX023980.D
Acq: 27 Aug 2021 09:42

Tgt Ion: 95 Resp: 139846
Ion Ratio Lower Upper
95 100
174 80.5 0.0 154.6
176 76.2 0.0 155.8








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.006 min
Lab File: VX023980.D
Acq: 27 Aug 2021 09:42

Instrument :
MSVOA_X
ClientSampleId :
VX0827WBL01

Tgt Ion:117 Resp: 268212
Ion Ratio Lower Upper
117 100
82 56.6 47.4 71.2
119 31.5 26.6 39.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX023980.D
Acq: 27 Aug 2021 09:42

Tgt Ion:152 Resp: 125089
Ion Ratio Lower Upper
152 100
115 59.4 41.1 123.3
150 154.2 0.0 349.0

