

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082722\
 Data File : VX030869.D
 Acq On : 27 Aug 2022 14:42
 Operator : JC/MD
 Sample : N4373-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT101D2-20220824

Quant Time: Aug 29 01:09:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	216988	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	373460	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	334377	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	149672	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	142376	55.782	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 111.560%			
35) Dibromofluoromethane	5.391	113	116960	50.958	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.920%			
50) Toluene-d8	8.653	98	437014	54.640	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 109.280%			
62) 4-Bromofluorobenzene	11.085	95	150213	50.446	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 100.900%			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	671	0.334	ug/l	83
9) 1,1,2-Trichlorotrifluo...	2.331	101	65191	30.851	ug/l	97
12) 1,1-Dichloroethene	2.319	96	18858	9.587	ug/l	98
24) 1,1-Dichloroethane	3.611	63	7039	1.699	ug/l	96
27) cis-1,2-Dichloroethene	4.495	96	7597	2.910	ug/l	96
29) Tetrahydrofuran	5.044	42	1367	1.102	ug/l #	76
30) Chloroform	5.092	83	5010	1.149	ug/l	93
38) Carbon Tetrachloride	5.690	117	4814	1.460	ug/l #	91
42) 1,2-Dichloroethane	6.098	62	1342	0.404	ug/l	97
44) Trichloroethene	7.135	130	3681990	1362.925	ug/l	97
55) 1,1,2-Trichloroethane	9.159	97	2615	1.008	ug/l	95
64) Tetrachloroethene	9.281	164	8106	3.077	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082722\
Data File : VX030869.D
Acq On : 27 Aug 2022 14:42
Operator : JC/MD
Sample : N4373-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT101D2-20220824

Quant Time: Aug 29 01:09:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

