

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082722\
 Data File : VX030871.D
 Acq On : 27 Aug 2022 15:29
 Operator : JC/MD
 Sample : N4389-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 265393-3

Quant Time: Aug 29 01:10:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	205926	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	348268	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	309089	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	131925	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	137986	56.966	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.940%
35) Dibromofluoromethane	5.391	113	110879	51.803	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.600%
50) Toluene-d8	8.653	98	418011	56.044	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	112.080%
62) 4-Bromofluorobenzene	11.085	95	138071	49.723	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082722\
Data File : VX030871.D
Acq On : 27 Aug 2022 15:29
Operator : JC/MD
Sample : N4389-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
265393-3

Quant Time: Aug 29 01:10:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

