

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082722\
 Data File : VX030877.D
 Acq On : 27 Aug 2022 17:50
 Operator : JC/MD
 Sample : N4387-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 251563-1

Quant Time: Aug 29 01:11:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	206127	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	356925	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	321610	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	137594	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	138432	57.095	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.180%
35) Dibromofluoromethane	5.391	113	113807	51.881	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.760%
50) Toluene-d8	8.653	98	424868	55.582	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	111.160%
62) 4-Bromofluorobenzene	11.085	95	141627	49.766	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.540%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	2936	2.620	ug/l #	86
44) Trichloroethene	7.129	130	940	0.364	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082722\
Data File : VX030877.D
Acq On : 27 Aug 2022 17:50
Operator : JC/MD
Sample : N4387-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
251563-1

Quant Time: Aug 29 01:11:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

