

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082722\
 Data File : VX030879.D
 Acq On : 27 Aug 2022 18:37
 Operator : JC/MD
 Sample : N4385-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 265751-1

Quant Time: Aug 29 01:11:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	202208	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	356974	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	324170	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	140182	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	139333	58.580	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	117.160%
35) Dibromofluoromethane	5.391	113	114491	52.186	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.380%
50) Toluene-d8	8.653	98	421895	55.186	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	110.380%
62) 4-Bromofluorobenzene	11.085	95	143759	50.508	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.020%
Target Compounds						Qvalue
44) Trichloroethene	7.141	130	831	0.322	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082722\
Data File : VX030879.D
Acq On : 27 Aug 2022 18:37
Operator : JC/MD
Sample : N4385-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
265751-1

Quant Time: Aug 29 01:11:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

