

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082722\
 Data File : VX030880.D
 Acq On : 27 Aug 2022 19:01
 Operator : JC/MD
 Sample : N4385-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 265751-2

Quant Time: Aug 29 01:11:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	205165	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	355485	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	314705	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	134696	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	140408	58.181	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 116.360%			
35) Dibromofluoromethane	5.391	113	112407	51.451	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 102.900%			
50) Toluene-d8	8.653	98	420318	55.210	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 110.420%			
62) 4-Bromofluorobenzene	11.085	95	141704	49.995	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 99.980%			
Target Compounds						Qvalue
44) Trichloroethene	7.135	130	685	0.266	ug/l	63

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082722\
Data File : VX030880.D
Acq On : 27 Aug 2022 19:01
Operator : JC/MD
Sample : N4385-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
265751-2

Quant Time: Aug 29 01:11:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

