

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082722\
 Data File : VX030883.D
 Acq On : 27 Aug 2022 20:12
 Operator : JC/MD
 Sample : N4297-19
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW8-MW01D3-20220821

Quant Time: Aug 29 01:12:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	202883	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	339039	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	295223	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	117510	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	137041	57.425	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.840%
35) Dibromofluoromethane	5.391	113	108438	52.042	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.080%
50) Toluene-d8	8.653	98	401222	55.258	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	110.520%
62) 4-Bromofluorobenzene	11.085	95	127067	47.005	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.020%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	2620	1.326	ug/l	95
29) Tetrahydrofuran	5.013	42	629027	542.535	ug/l	95
44) Trichloroethene	7.128	130	1252	0.510	ug/l	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082722\
Data File : VX030883.D
Acq On : 27 Aug 2022 20:12
Operator : JC/MD
Sample : N4297-19
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW8-MW01D3-20220821

Quant Time: Aug 29 01:12:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

