

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082722\
 Data File : VX030885.D
 Acq On : 27 Aug 2022 20:59
 Operator : JC/MD
 Sample : N4297-21
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW9-MW01D3-20220821

Quant Time: Aug 29 01:12:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	196198	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	339487	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	310823	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	130814	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	137507	59.583	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	119.160%
35) Dibromofluoromethane	5.391	113	109994	52.719	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.440%
50) Toluene-d8	8.653	98	407573	56.058	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	112.120%
62) 4-Bromofluorobenzene	11.085	95	134773	49.790	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.580%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.337	101	723	0.378	ug/l	# 84
16) Acetone	2.386	43	2145	2.011	ug/l	98
44) Trichloroethene	7.135	130	561	0.228	ug/l	64

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082722\
Data File : VX030885.D
Acq On : 27 Aug 2022 20:59
Operator : JC/MD
Sample : N4297-21
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW9-MW01D3-20220821

Quant Time: Aug 29 01:12:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

