

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082722\
 Data File : VX030886.D
 Acq On : 27 Aug 2022 21:22
 Operator : JC/MD
 Sample : N4297-22
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW9-MW01D2-20220821

Quant Time: Aug 29 01:12:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	198485	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	342791	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	309600	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	130621	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	135283	57.944	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	115.880%
35) Dibromofluoromethane	5.391	113	110908	52.644	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.280%
50) Toluene-d8	8.653	98	406633	55.390	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	110.780%
62) 4-Bromofluorobenzene	11.085	95	134849	49.338	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.680%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	648	0.335	ug/l #	69
16) Acetone	2.386	43	5700	5.283	ug/l	99
29) Tetrahydrofuran	5.026	42	21249	18.733	ug/l	92
44) Trichloroethene	7.129	130	611	0.246	ug/l	64

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082722\
Data File : VX030886.D
Acq On : 27 Aug 2022 21:22
Operator : JC/MD
Sample : N4297-22
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW9-MW01D2-20220821

Quant Time: Aug 29 01:12:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

