

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082722\
 Data File : VX030887.D
 Acq On : 27 Aug 2022 21:46
 Operator : JC/MD
 Sample : N4297-23
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW9-MW01S-20220821

Quant Time: Aug 29 01:12:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	192465	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	327979	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	292882	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	122364	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	131733	58.188	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	116.380%
35) Dibromofluoromethane	5.397	113	106056	52.615	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.220%
50) Toluene-d8	8.653	98	390055	55.531	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	111.060%
62) 4-Bromofluorobenzene	11.085	95	126172	48.248	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.500%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.325	101	407	0.217	ug/l	# 75
16) Acetone	2.386	43	2062	1.971	ug/l	91
29) Tetrahydrofuran	5.026	42	42750	38.868	ug/l	94
44) Trichloroethene	7.135	130	556	0.234	ug/l	65

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082722\
Data File : VX030887.D
Acq On : 27 Aug 2022 21:46
Operator : JC/MD
Sample : N4297-23
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW9-MW01S-20220821

Quant Time: Aug 29 01:12:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

