

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082722\
 Data File : VX030894.D
 Acq On : 28 Aug 2022 01:40
 Operator : JC/MD
 Sample : N4360-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB02-20220823

Quant Time: Aug 29 01:14:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) Pentafluorobenzene		5.562	168	191813	50.000	ug/l	0.00
34) 1,4-Difluorobenzene		6.769	114	327365	50.000	ug/l	0.00
63) Chlorobenzene-d5		10.055	117	295364	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4		12.024	152	121952	50.000	ug/l	0.00
System Monitoring Compounds							
33) 1,2-Dichloroethane-d4		5.964	65	137809	61.079	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	122.160%	
35) Dibromofluoromethane		5.391	113	107428	53.395	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.800%	
50) Toluene-d8		8.653	98	398445	56.832	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	113.660%#	
62) 4-Bromofluorobenzene		11.085	95	130406	49.961	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.920%	
Target Compounds							
							Qvalue
16) Acetone		2.386	43	4054	3.888	ug/l	99
25) 2-Butanone		4.587	43	3284	1.955	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082722\
Data File : VX030894.D
Acq On : 28 Aug 2022 01:40
Operator : JC/MD
Sample : N4360-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB02-20220823

Quant Time: Aug 29 01:14:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

