

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082722\
 Data File : VX030898.D
 Acq On : 28 Aug 2022 03:14
 Operator : JC/MD
 Sample : N4360-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT172S1-20220823

Quant Time: Aug 29 01:15:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	181278	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	320887	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	285384	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	116941	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	136390	63.963	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 127.920%#	
35) Dibromofluoromethane	5.397	113	104651	53.065	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.140%	
50) Toluene-d8	8.653	98	383726	55.838	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 111.680%	
62) 4-Bromofluorobenzene	11.085	95	125752	49.151	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 98.300%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	2961	1.677	ug/l #	91
12) 1,1-Dichloroethene	2.313	96	527	0.321	ug/l #	60
16) Acetone	2.386	43	1848	1.875	ug/l	92
19) Methyl tert-butyl Ether	3.117	73	1829	0.321	ug/l #	89
27) cis-1,2-Dichloroethene	4.501	96	850	0.390	ug/l	85
30) Chloroform	5.111	83	1701	0.467	ug/l #	83
44) Trichloroethene	7.129	130	88409	38.087	ug/l	99
64) Tetrachloroethene	9.281	164	2687	1.195	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082722\
Data File : VX030898.D
Acq On : 28 Aug 2022 03:14
Operator : JC/MD
Sample : N4360-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT172S1-20220823

Quant Time: Aug 29 01:15:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

