

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082722\
 Data File : VX030899.D
 Acq On : 28 Aug 2022 03:37
 Operator : JC/MD
 Sample : N4360-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE125D2-20220824

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 08/29/2022
 Supervised By :Mahesh Dadoda 08/29/2022

Quant Time: Aug 29 01:15:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	183566	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	321924	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	284750	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	119313	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	135132	62.584	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 125.160%#	
35) Dibromofluoromethane	5.391	113	103433	52.279	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.560%	
50) Toluene-d8	8.653	98	384380	55.753	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 111.500%	
62) 4-Bromofluorobenzene	11.085	95	123844	48.249	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 96.500%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	1136	0.668	ug/l	96
7) Trichlorofluoromethane	1.886	101	917	0.326	ug/l #	76
9) 1,1,2-Trichlorotrifluo...	2.331	101	32341	18.092	ug/l	98
12) 1,1-Dichloroethene	2.319	96	8770	5.271	ug/l	85
27) cis-1,2-Dichloroethene	4.495	96	11378	5.153	ug/l	98
30) Chloroform	5.117	83	2024m	0.549	ug/l	
38) Carbon Tetrachloride	5.690	117	1915	0.674	ug/l #	66
42) 1,2-Dichloroethane	6.092	62	1059	0.370	ug/l	86
44) Trichloroethene	7.129	130	763758	327.971	ug/l	92
55) 1,1,2-Trichloroethane	9.153	97	872	0.390	ug/l #	79
64) Tetrachloroethene	9.275	164	10483	4.674	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082722\
Data File : VX030899.D
Acq On : 28 Aug 2022 03:37
Operator : JC/MD
Sample : N4360-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE125D2-20220824

Quant Time: Aug 29 01:15:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 08/29/2022
Supervised By :Mahesh Dadoda 08/29/2022

