

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082722\
 Data File : VX030900.D
 Acq On : 28 Aug 2022 04:01
 Operator : JC/MD
 Sample : N4360-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GM38-RW3-MW1-20220824

Quant Time: Aug 29 01:16:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	184713	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	323266	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	288461	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	120670	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	137786	63.416	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 126.840%#	
35) Dibromofluoromethane	5.391	113	104785	52.742	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.480%	
50) Toluene-d8	8.653	98	389896	56.318	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 112.640%	
62) 4-Bromofluorobenzene	11.079	95	130592	50.667	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.340%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	487	0.271	ug/l	89
44) Trichloroethene	7.129	130	48917	20.919	ug/l	89
64) Tetrachloroethene	9.275	164	3058	1.346	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082722\
Data File : VX030900.D
Acq On : 28 Aug 2022 04:01
Operator : JC/MD
Sample : N4360-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GM38-RW3-MW1-20220824

Quant Time: Aug 29 01:16:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

