

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082722\
 Data File : VX030902.D
 Acq On : 28 Aug 2022 04:48
 Operator : JC/MD
 Sample : N4360-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE135D1-20220824

Quant Time: Aug 29 01:16:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	175317	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	303655	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	275553	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	116431	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	133962	64.961	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 129.920%#	
35) Dibromofluoromethane	5.391	113	97767	52.388	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.780%	
50) Toluene-d8	8.653	98	368837	56.717	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 113.440%#	
62) 4-Bromofluorobenzene	11.085	95	123714	51.098	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.200%	
Target Compounds						Qvalue
30) Chloroform	5.099	83	6169	1.751	ug/l	92
44) Trichloroethene	7.129	130	1150	0.524	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082722\
Data File : VX030902.D
Acq On : 28 Aug 2022 04:48
Operator : JC/MD
Sample : N4360-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE135D1-20220824

Quant Time: Aug 29 01:16:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

