

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082722\
 Data File : VX030903.D
 Acq On : 28 Aug 2022 05:11
 Operator : JC/MD
 Sample : N4360-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT101D-20220824

Quant Time: Aug 29 01:16:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	167905	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	300145	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	264693	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	110039	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	132018	66.844	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 133.680%#	
35) Dibromofluoromethane	5.397	113	96646	52.393	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.780%	
50) Toluene-d8	8.653	98	353424	54.982	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 109.960%	
62) 4-Bromofluorobenzene	11.085	95	118638	49.574	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 99.140%	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	2666	1.713	ug/l	97
9) 1,1,2-Trichlorotrifluo...	2.331	101	27440	16.782	ug/l	96
12) 1,1-Dichloroethene	2.319	96	4823	3.169	ug/l	84
24) 1,1-Dichloroethane	3.617	63	3358	1.047	ug/l	97
27) cis-1,2-Dichloroethene	4.501	96	8107	4.014	ug/l	97
30) Chloroform	5.099	83	1965	0.582	ug/l	99
44) Trichloroethene	7.129	130	258961	119.272	ug/l	90
64) Tetrachloroethene	9.281	164	1567	0.752	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082722\
Data File : VX030903.D
Acq On : 28 Aug 2022 05:11
Operator : JC/MD
Sample : N4360-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT101D-20220824

Quant Time: Aug 29 01:16:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

