

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082722\
 Data File : VX030905.D
 Acq On : 28 Aug 2022 05:58
 Operator : JC/MD
 Sample : N4360-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE135D3-20220824

Quant Time: Aug 29 01:17:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	149846	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	260582	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	225911	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85944	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	121836	69.123	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 138.240%#		
35) Dibromofluoromethane	5.397	113	86522	54.026	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 108.060%		
50) Toluene-d8	8.653	98	311196	55.763	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 111.520%		
62) 4-Bromofluorobenzene	11.085	95	98366	47.344	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 94.680%		
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	21008	14.396	ug/l	95
12) 1,1-Dichloroethene	2.319	96	1214	0.894	ug/l	84
27) cis-1,2-Dichloroethene	4.495	96	10495	5.822	ug/l	91
44) Trichloroethene	7.129	130	46537	24.688	ug/l	89
45) 1,2-Dichloropropane	7.434	63	7917	4.386	ug/l	91
64) Tetrachloroethene	9.275	164	61393	34.499	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082722\
Data File : VX030905.D
Acq On : 28 Aug 2022 05:58
Operator : JC/MD
Sample : N4360-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE135D3-20220824

Quant Time: Aug 29 01:17:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

