

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX082722\
 Data File : VX030864.D
 Acq On : 27 Aug 2022 12:12
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 29 01:08:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	89	0.00
2 T	Dichlorodifluoromethane	50.000	47.886	4.2	90	0.00
3 P	Chloromethane	50.000	50.408	-0.8	91	0.00
4 C	Vinyl Chloride	50.000	49.972	0.1#	90	0.00
5 T	Bromomethane	50.000	50.597	-1.2	104	0.00
6 T	Chloroethane	50.000	53.762	-7.5	96	0.00
7 T	Trichlorofluoromethane	50.000	47.716	4.6	95	0.00
8 T	Diethyl Ether	50.000	50.444	-0.9	99	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	46.221	7.6	90	0.00
10 T	Methyl Iodide	50.000	43.131	13.7	81	0.00
11 T	Tert butyl alcohol	250.000	240.181	3.9	98	0.00
12 CM	1,1-Dichloroethene	50.000	43.478	13.0#	87	0.00
13 T	Acrolein	250.000	258.458	-3.4	97	0.00
14 T	Allyl chloride	50.000	47.956	4.1	94	0.00
15 T	Acrylonitrile	250.000	247.677	0.9	93	0.00
16 T	Acetone	250.000	235.377	5.8	90	0.00
17 T	Carbon Disulfide	50.000	45.350	9.3	83	0.00
18 T	Methyl Acetate	50.000	49.260	1.5	95	0.00
19 T	Methyl tert-butyl Ether	50.000	50.192	-0.4	93	0.00
20 T	Methylene Chloride	50.000	50.489	-1.0	91	0.00
21 T	trans-1,2-Dichloroethene	50.000	44.713	10.6	89	0.00
22 T	Diisopropyl ether	50.000	51.732	-3.5	98	0.00
23 T	Vinyl Acetate	250.000	261.250	-4.5	97	0.00
24 P	1,1-Dichloroethane	50.000	47.402	5.2	92	0.00
25 T	2-Butanone	250.000	246.586	1.4	93	0.00
26 T	2,2-Dichloropropane	50.000	50.220	-0.4	93	0.00
27 T	cis-1,2-Dichloroethene	50.000	45.740	8.5	90	0.00
28 T	Bromochloromethane	50.000	49.936	0.1	96	0.00
29 T	Tetrahydrofuran	250.000	244.551	2.2	94	0.00
30 C	Chloroform	50.000	48.275	3.5#	96	0.00
31 T	Cyclohexane	50.000	46.596	6.8	88	0.00
32 T	1,1,1-Trichloroethane	50.000	49.420	1.2	93	0.00
33 S	1,2-Dichloroethane-d4	50.000	48.133	3.7	97	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	90	0.00
35 S	Dibromofluoromethane	50.000	47.266	5.5	91	0.00
36 T	1,1-Dichloropropene	50.000	45.291	9.4	90	0.00
37 T	Ethyl Acetate	50.000	49.793	0.4	97	0.00
38 T	Carbon Tetrachloride	50.000	47.447	5.1	92	0.00
39 T	Methylcyclohexane	50.000	44.406	11.2	86	0.00
40 TM	Benzene	50.000	47.712	4.6	93	0.00
41 T	Methacrylonitrile	50.000	51.168	-2.3	96	0.00
42 TM	1,2-Dichloroethane	50.000	51.036	-2.1	99	-0.01
43 T	Isopropyl Acetate	50.000	52.443	-4.9	98	0.00
44 TM	Trichloroethene	50.000	45.376	9.2	88	0.00
45 C	1,2-Dichloropropane	50.000	48.980	2.0#	93	0.00
46 T	Dibromomethane	50.000	46.259	7.5	90	0.00
47 T	Bromodichloromethane	50.000	50.677	-1.4	94	0.00
48 T	Methyl methacrylate	50.000	52.949	-5.9	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082722\
 Data File : VX030864.D
 Acq On : 27 Aug 2022 12:12
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 29 01:08:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	975.760	2.4	95	0.00
50 S	Toluene-d8	50.000	48.295	3.4	93	0.00
51 T	4-Methyl-2-Pentanone	250.000	263.861	-5.5	98	0.00
52 CM	Toluene	50.000	48.684	2.6#	92	0.00
53 T	t-1,3-Dichloropropene	50.000	52.815	-5.6	95	0.00
54 T	cis-1,3-Dichloropropene	50.000	51.415	-2.8	93	0.00
55 T	1,1,2-Trichloroethane	50.000	50.226	-0.5	94	0.00
56 T	Ethyl methacrylate	50.000	53.260	-6.5	94	0.00
57 T	1,3-Dichloropropane	50.000	50.292	-0.6	95	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	267.894	-7.2	94	0.00
59 T	2-Hexanone	250.000	267.263	-6.9	97	0.00
60 T	Dibromochloromethane	50.000	49.565	0.9	91	0.00
61 T	1,2-Dibromoethane	50.000	49.680	0.6	92	0.00
62 S	4-Bromofluorobenzene	50.000	51.051	-2.1	95	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	93	0.00
64 T	Tetrachloroethene	50.000	46.061	7.9	91	0.00
65 PM	Chlorobenzene	50.000	47.191	5.6	94	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	48.104	3.8	92	0.00
67 C	Ethyl Benzene	50.000	49.097	1.8#	95	0.00
68 T	m/p-Xylenes	100.000	98.123	1.9	94	0.00
69 T	o-Xylene	50.000	48.830	2.3	93	0.00
70 T	Styrene	50.000	50.503	-1.0	93	0.00
71 P	Bromoform	50.000	46.166	7.7	87	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	90	0.00
73 T	Isopropylbenzene	50.000	48.683	2.6	94	0.00
74 T	N-ethyl acetate	50.000	54.556	-9.1	99	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	47.466	5.1	94	0.00
76 T	1,2,3-Trichloropropane	50.000	48.196	3.6	84	0.00
77 T	Bromobenzene	50.000	47.105	5.8	94	0.00
78 T	n-propylbenzene	50.000	50.330	-0.7	95	0.00
79 T	2-Chlorotoluene	50.000	48.576	2.8	96	0.00
80 T	1,3,5-Trimethylbenzene	50.000	49.446	1.1	95	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	50.649	-1.3	92	0.00
82 T	4-Chlorotoluene	50.000	49.843	0.3	95	0.00
83 T	tert-Butylbenzene	50.000	49.728	0.5	92	0.00
84 T	1,2,4-Trimethylbenzene	50.000	50.491	-1.0	93	0.00
85 T	sec-Butylbenzene	50.000	50.614	-1.2	93	0.00
86 T	p-Isopropyltoluene	50.000	50.768	-1.5	93	0.00
87 T	1,3-Dichlorobenzene	50.000	47.649	4.7	86	0.00
88 T	1,4-Dichlorobenzene	50.000	47.546	4.9	89	0.00
89 T	n-Butylbenzene	50.000	50.154	-0.3	87	0.00
90 T	Hexachloroethane	50.000	45.034	9.9	84	0.00
91 T	1,2-Dichlorobenzene	50.000	46.281	7.4	88	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	43.741	12.5	81	0.00
93 T	1,2,4-Trichlorobenzene	50.000	46.151	7.7	82	0.00
94 T	Hexachlorobutadiene	50.000	47.079	5.8	85	0.00
95 T	Naphthalene	50.000	48.174	3.7	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082722\
Data File : VX030864.D
Acq On : 27 Aug 2022 12:12
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Aug 29 01:08:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	45.355	9.3	79	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6