

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082722\
 Data File : VX030875.D
 Acq On : 27 Aug 2022 17:03
 Operator : JC/MD
 Sample : N4383-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 252672-1

Quant Time: Aug 29 01:10:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	5.556	168	207187	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	357304	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	323597	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	137440	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	140201	57.528	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	115.060%
35) Dibromofluoromethane	5.391	113	112362	51.168	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.340%
50) Toluene-d8	8.653	98	422562	55.222	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	110.440%
62) 4-Bromofluorobenzene	11.085	95	140782	49.417	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.840%
Target Compounds						
					Qvalue	
44) Trichloroethene	7.135	130	1606	0.621	ug/l	74


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082722\
Data File : VX030875.D
Acq On : 27 Aug 2022 17:03
Operator : JC/MD
Sample : N4383-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
252672-1

Quant Time: Aug 29 01:10:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX030875.D
Acq: 27 Aug 2022 17:03




Instrument :
MSVOA_X
ClientSampleId :
252672-1


Tgt Ion:168 Resp: 207187
Ion Ratio Lower Upper
168 100
99 58.1 44.0 66.0



#33
1,2-Dichloroethane-d4
Concen: 57.528 ug/l
RT: 5.964 min Scan# 800
Delta R.T. 0.006 min
Lab File: VX030875.D
Acq: 27 Aug 2022 17:03

Tgt Ion: 65 Resp: 140201
Ion Ratio Lower Upper
65 100
67 53.1 0.0 109.6









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.769 min Scan# 911
Delta R.T. 0.000 min
Lab File: VX030875.D
Acq: 27 Aug 2022 17:03




Instrument :
MSVOA_X
ClientSampleId :
252672-1


Tgt Ion:114 Resp: 357304
Ion Ratio Lower Upper
114 100
63 20.8 0.0 38.4
88 15.3 0.0 31.8



#35
Dibromofluoromethane
Concen: 51.168 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.000 min
Lab File: VX030875.D
Acq: 27 Aug 2022 17:03

Tgt Ion:113 Resp: 112362
Ion Ratio Lower Upper
113 100
111 102.6 83.2 124.8
192 18.1 14.8 22.2









#44
Trichloroethene
Concen: 0.621 ug/l
RT: 7.135 min Scan# 991
Delta R.T. 0.006 min
Lab File: VX030875.D
Acq: 27 Aug 2022 17:03




Instrument :
MSVOA_X
ClientSampleId :
252672-1


Tgt Ion:130 Resp: 1606
Ion Ratio Lower Upper
130 100
95 72.0 0.0 194.8



#50
Toluene-d8
Concen: 55.222 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX030875.D
Acq: 27 Aug 2022 17:03

Tgt Ion: 98 Resp: 422562
Ion Ratio Lower Upper
98 100
100 63.5 52.3 78.5









#62
4-Bromofluorobenzene
Concen: 49.417 ug/l
RT: 11.085 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX030875.D
Acq: 27 Aug 2022 17:03




Instrument :
MSVOA_X
ClientSampleId :
252672-1


Tgt Ion: 95 Resp: 140782
Ion Ratio Lower Upper
95 100
174 77.0 0.0 149.0
176 73.9 0.0 146.0



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX030875.D
Acq: 27 Aug 2022 17:03

Tgt Ion: 117 Resp: 323597
Ion Ratio Lower Upper
117 100
82 57.1 42.7 64.1
119 30.4 25.4 38.2





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX030875.D


Acq: 27 Aug 2022 17:03

Instrument :

MSVOA_X

ClientSampleId :

252672-1



Tgt Ion:152 Resp: 137440

Ion Ratio Lower Upper

152 100

115 59.7 40.5 121.5

150 157.9 0.0 347.0

