

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082725\
 Data File : VX047536.D
 Acq On : 27 Aug 2025 13:14
 Operator : JC/MD
 Sample : VX0827WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0827WBL01

Quant Time: Aug 28 01:18:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 18 04:18:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	194583	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	393301	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	391930	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	194932	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	146855	53.926	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	107.860%
35) Dibromofluoromethane	5.397	113	123885	48.534	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.060%
50) Toluene-d8	8.647	98	433620	47.528	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.060%
62) 4-Bromofluorobenzene	11.079	95	182336	54.158	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082725\
Data File : VX047536.D
Acq On : 27 Aug 2025 13:14
Operator : JC/MD
Sample : VX0827WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0827WBL01

Quant Time: Aug 28 01:18:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
Quant Title : SW846 8260
QLast Update : Mon Aug 18 04:18:28 2025
Response via : Initial Calibration

