

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082818\
 Data File : VX004274.D
 Acq On : 28 Aug 2018 17:57
 Operator : JC/MD
 Sample : J4662-01DL 10X
 Misc : 5.31g/5.0mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EP1-D7-GRABDL

Quant Time: Aug 29 00:54:49 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 23 05:39:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	290031	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	447072	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	447188	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	286195	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	192857	51.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.04%	
35) Dibromofluoromethane	5.50	113	166221	49.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.96%	
50) Toluene-d8	8.71	98	635999	50.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.32%	
62) 4-Bromofluorobenzene	11.14	95	300945	64.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	128.50%	
Target Compounds						
73) Isopropylbenzene	11.02	105	85360	4.54	ug/l	99
78) n-propylbenzene	11.36	91	349134	16.15	ug/l	97
85) sec-Butylbenzene	11.95	105	461786	24.84	ug/l	99
86) p-Isopropyltoluene	12.07	119	55121	3.31	ug/l	97
89) n-Butylbenzene	12.39	91	643311	47.48	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082818\
Data File : VX004274.D
Acq On : 28 Aug 2018 17:57
Operator : JC/MD
Sample : J4662-01DL 10X
Misc : 5.31g/5.0mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EP1-D7-GRABDL

Quant Time: Aug 29 00:54:49 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 23 05:39:44 2018
Response via : Initial Calibration

