

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082820\
 Data File : VX018117.D
 Acq On : 28 Aug 2020 13:58
 Operator : JC/SP
 Sample : L3827-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 789UMR-DSP01-01

Quant Time: Aug 29 03:45:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	624361	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	1031075	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	974976	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	440714	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	431044	49.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.46%	
35) Dibromofluoromethane	5.47	113	345042	49.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.28%	
50) Toluene-d8	8.70	98	1280712	48.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.94%	
62) 4-Bromofluorobenzene	11.12	95	467061	45.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.78%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	3.02	59	11367	6.745	ug/l #	86
16) Acetone	2.42	43	176803	52.348	ug/l	97
18) Methyl Acetate	2.76	43	10356	1.271	ug/l #	87
20) Methylene Chloride	2.84	84	41807	5.003	ug/l	94
25) 2-Butanone	4.65	43	45904	8.270	ug/l	98
43) Isopropyl Acetate	6.42	43	32835	1.840	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082820\
Data File : VX018117.D
Acq On : 28 Aug 2020 13:58
Operator : JC/SP
Sample : L3827-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
789UMR-DSP01-01

Quant Time: Aug 29 03:45:01 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

