

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082823\
 Data File : VX037286.D
 Acq On : 28 Aug 2023 11:43
 Operator : JC/MD
 Sample : VX0828MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0828MBL01

Quant Time: Aug 29 00:55:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	149828	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	241154	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	211763	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95725	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	90061	41.694	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	83.380%
35) Dibromofluoromethane	5.385	113	74865	46.875	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.760%
50) Toluene-d8	8.647	98	282164	48.705	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.420%
62) 4-Bromofluorobenzene	11.079	95	102533	45.036	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082823\
Data File : VX037286.D
Acq On : 28 Aug 2023 11:43
Operator : JC/MD
Sample : VX0828MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0828MBL01

Quant Time: Aug 29 00:55:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

