

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082823\
 Data File : VX037290.D
 Acq On : 28 Aug 2023 13:22
 Operator : JC/MD
 Sample : 04095-18DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT101D1-20230818DL

Quant Time: Aug 29 00:57:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	148012	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	241335	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	213674	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97895	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	87818	41.155	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	82.300%
35) Dibromofluoromethane	5.385	113	73755	46.146	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.300%
50) Toluene-d8	8.647	98	280786	48.431	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.860%
62) 4-Bromofluorobenzene	11.079	95	103813	45.564	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.120%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.325	101	1747	1.117	ug/l	94
44) Trichloroethene	7.123	130	61269	33.178	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082823\
Data File : VX037290.D
Acq On : 28 Aug 2023 13:22
Operator : JC/MD
Sample : 04095-18DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT101D1-20230818DL

Quant Time: Aug 29 00:57:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

