

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082823\
 Data File : VX037292.D
 Acq On : 28 Aug 2023 14:10
 Operator : JC/MD
 Sample : 04095-13 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT101D2-20230818

Quant Time: Aug 29 00:58:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	148471	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	244264	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	216819	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96981	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	89868	41.985	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	83.980%
35) Dibromofluoromethane	5.385	113	74297	45.927	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.860%
50) Toluene-d8	8.646	98	285418	48.640	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.280%
62) 4-Bromofluorobenzene	11.079	95	105333	45.677	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.360%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.324	101	2286	1.457	ug/l	96
12) 1,1-Dichloroethene	2.312	96	896	0.586	ug/l	95
20) Methylene Chloride	2.788	84	1249	0.667	ug/l #	71
44) Trichloroethene	7.122	130	148402	79.398	ug/l	96
64) Tetrachloroethene	9.274	164	565	0.357	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082823\
Data File : VX037292.D
Acq On : 28 Aug 2023 14:10
Operator : JC/MD
Sample : 04095-13 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT101D2-20230818

Quant Time: Aug 29 00:58:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

