

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082823\
 Data File : VX037295.D
 Acq On : 28 Aug 2023 15:23
 Operator : JC/MD
 Sample : 04126-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB05-20230819

Quant Time: Aug 29 00:59:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	145209	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	237250	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	205125	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	86612	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	89595	42.798	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.600%
35) Dibromofluoromethane	5.385	113	72753	46.303	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.600%
50) Toluene-d8	8.647	98	276136	48.449	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.900%
62) 4-Bromofluorobenzene	11.079	95	96836	43.234	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082823\
Data File : VX037295.D
Acq On : 28 Aug 2023 15:23
Operator : JC/MD
Sample : 04126-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB05-20230819

Quant Time: Aug 29 00:59:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

