

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082823\
 Data File : VX037297.D
 Acq On : 28 Aug 2023 16:12
 Operator : JC/MD
 Sample : 04126-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE123D3-20230819

Quant Time: Aug 29 00:59:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	143709	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	237204	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	209593	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92959	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	89170	43.039	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.080%
35) Dibromofluoromethane	5.385	113	72391	46.081	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.160%
50) Toluene-d8	8.647	98	275874	48.412	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.820%
62) 4-Bromofluorobenzene	11.079	95	100677	44.957	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.920%
Target Compounds						
					Qvalue	
44) Trichloroethene	7.129	130	3159	1.740	ug/l	86
64) Tetrachloroethene	9.275	164	1730	1.131	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082823\
Data File : VX037297.D
Acq On : 28 Aug 2023 16:12
Operator : JC/MD
Sample : 04126-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE123D3-20230819

Quant Time: Aug 29 00:59:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

