

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082823\
 Data File : VX037298.D
 Acq On : 28 Aug 2023 16:36
 Operator : JC/MD
 Sample : 04126-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW202D-20230819

Quant Time: Aug 29 01:00:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	139778	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	228192	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	198168	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	86610	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	84640	42.002	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	84.000%
35) Dibromofluoromethane	5.385	113	69545	46.018	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.040%
50) Toluene-d8	8.647	98	266188	48.557	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.120%
62) 4-Bromofluorobenzene	11.079	95	96362	44.730	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.460%
Target Compounds						
					Qvalue	
7) Trichlorofluoromethane	1.879	101	1878	0.741	ug/l	96
12) 1,1-Dichloroethene	2.312	96	751	0.522	ug/l	89
19) Methyl tert-butyl Ether	3.117	73	2877	0.527	ug/l	100
24) 1,1-Dichloroethane	3.605	63	4116	1.436	ug/l	93
27) cis-1,2-Dichloroethene	4.489	96	6680	3.464	ug/l	83
44) Trichloroethene	7.128	130	31496	18.038	ug/l	96
64) Tetrachloroethene	9.274	164	8376	5.794	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082823\
Data File : VX037298.D
Acq On : 28 Aug 2023 16:36
Operator : JC/MD
Sample : 04126-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW202D-20230819

Quant Time: Aug 29 01:00:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

