

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082823\
 Data File : VX037300.D
 Acq On : 28 Aug 2023 17:25
 Operator : JC/MD
 Sample : 04126-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP02-20230819

Quant Time: Aug 29 01:01:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	142386	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	229500	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	198356	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85932	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	87442	42.598	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.200%
35) Dibromofluoromethane	5.385	113	70980	46.700	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.400%
50) Toluene-d8	8.647	98	270588	49.079	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.160%
62) 4-Bromofluorobenzene	11.079	95	94054	43.410	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.820%
Target Compounds						
					Qvalue	
7) Trichlorofluoromethane	1.880	101	1923	0.745	ug/l	90
12) 1,1-Dichloroethene	2.313	96	707	0.482	ug/l	87
19) Methyl tert-butyl Ether	3.117	73	2820	0.507	ug/l	98
24) 1,1-Dichloroethane	3.605	63	4004	1.372	ug/l	99
27) cis-1,2-Dichloroethene	4.489	96	6962	3.544	ug/l	79
44) Trichloroethene	7.123	130	31494	17.934	ug/l	97
64) Tetrachloroethene	9.275	164	8233	5.690	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082823\
Data File : VX037300.D
Acq On : 28 Aug 2023 17:25
Operator : JC/MD
Sample : 04126-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP02-20230819

Quant Time: Aug 29 01:01:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

