

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082823\
 Data File : VX037301.D
 Acq On : 28 Aug 2023 17:49
 Operator : JC/MD
 Sample : 04126-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP07-20230819

Quant Time: Aug 29 01:01:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	139431	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	227313	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	201188	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85902	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	84853	42.212	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	84.420%
35) Dibromofluoromethane	5.385	113	69613	46.241	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.480%
50) Toluene-d8	8.647	98	265694	48.655	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.300%
62) 4-Bromofluorobenzene	11.079	95	93966	43.786	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.580%
Target Compounds						
					Qvalue	
44) Trichloroethene	7.129	130	3322	1.910	ug/l	80
52) Toluene	8.720	92	941	0.232	ug/l	97
64) Tetrachloroethene	9.275	164	1895	1.291	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082823\
Data File : VX037301.D
Acq On : 28 Aug 2023 17:49
Operator : JC/MD
Sample : 04126-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP07-20230819

Quant Time: Aug 29 01:01:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

