

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082823\
 Data File : VX037303.D
 Acq On : 28 Aug 2023 18:37
 Operator : JC/MD
 Sample : 04126-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW201D1-20230819

Quant Time: Aug 29 01:02:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	140522	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	230698	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	201833	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	88747	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	85299	42.105	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	84.200%
35) Dibromofluoromethane	5.385	113	70026	45.833	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.660%
50) Toluene-d8	8.647	98	271544	48.996	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.000%
62) 4-Bromofluorobenzene	11.079	95	97344	44.695	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.380%
Target Compounds						
					Qvalue	
7) Trichlorofluoromethane	1.874	101	994	0.390	ug/l	99
44) Trichloroethene	7.129	130	1726	0.978	ug/l	97
64) Tetrachloroethene	9.275	164	1591	1.081	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082823\
Data File : VX037303.D
Acq On : 28 Aug 2023 18:37
Operator : JC/MD
Sample : 04126-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW201D1-20230819

Quant Time: Aug 29 01:02:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

