

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082823\
 Data File : VX037313.D
 Acq On : 29 Aug 2023 02:21
 Operator : JC/MD
 Sample : 04126-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW203D-20230820

Quant Time: Aug 29 05:31:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	149526	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	245927	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	216215	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96783	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	91981	42.669	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.340%
35) Dibromofluoromethane	5.385	113	75979	46.650	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.300%
50) Toluene-d8	8.647	98	288702	48.867	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.740%
62) 4-Bromofluorobenzene	11.079	95	104628	45.064	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.120%
Target Compounds						
					Qvalue	
24) 1,1-Dichloroethane	3.611	63	4593	1.498	ug/l	95
27) cis-1,2-Dichloroethene	4.489	96	5757	2.791	ug/l	78
44) Trichloroethene	7.123	130	25602	13.605	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082823\
Data File : VX037313.D
Acq On : 29 Aug 2023 02:21
Operator : JC/MD
Sample : 04126-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW203D-20230820

Quant Time: Aug 29 05:31:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

