

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082823\
 Data File : VX037318.D
 Acq On : 29 Aug 2023 04:21
 Operator : JC/MD
 Sample : 04126-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE131D2-20230821

Quant Time: Aug 29 07:20:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	152150	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	249525	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	222105	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	102394	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	92455	42.149	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	84.300%
35) Dibromofluoromethane	5.385	113	75949	45.959	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.920%
50) Toluene-d8	8.647	98	292275	48.758	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.520%
62) 4-Bromofluorobenzene	11.079	95	108627	46.112	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.220%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.325	101	191256	118.914	ug/l	96
12) 1,1-Dichloroethene	2.319	96	3118	1.991	ug/l #	78
22) Diisopropyl ether	3.764	45	1504	0.302	ug/l #	66
27) cis-1,2-Dichloroethene	4.489	96	8490	4.045	ug/l	79
44) Trichloroethene	7.123	130	265568	139.089	ug/l	98
64) Tetrachloroethene	9.275	164	25237	15.576	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082823\
Data File : VX037318.D
Acq On : 29 Aug 2023 04:21
Operator : JC/MD
Sample : 04126-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE131D2-20230821

Quant Time: Aug 29 07:20:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

