

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082823\
 Data File : VX037319.D
 Acq On : 29 Aug 2023 04:45
 Operator : JC/MD
 Sample : 04126-17DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE131D2-20230821DL

Quant Time: Aug 29 07:20:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	150071	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	248240	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	217706	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	98613	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	92745	42.867	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.740%
35) Dibromofluoromethane	5.385	113	74979	45.607	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.220%
50) Toluene-d8	8.647	98	287961	48.287	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.580%
62) 4-Bromofluorobenzene	11.079	95	105210	44.893	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.780%
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.325	101	35434	22.336	ug/l	97
27) cis-1,2-Dichloroethene	4.495	96	1539	0.743	ug/l	81
44) Trichloroethene	7.123	130	50605	26.641	ug/l	96
64) Tetrachloroethene	9.275	164	4874	3.069	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082823\
Data File : VX037319.D
Acq On : 29 Aug 2023 04:45
Operator : JC/MD
Sample : 04126-17DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE131D2-20230821DL

Quant Time: Aug 29 07:20:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

