

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082823\
 Data File : VX037322.D
 Acq On : 29 Aug 2023 05:58
 Operator : JC/MD
 Sample : 04126-21
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D4-20230821

Quant Time: Aug 29 07:21:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	156915	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	259013	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	228772	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	101348	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	97461	43.082	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	86.160%		
35) Dibromofluoromethane	5.385	113	78710	45.885	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	91.760%		
50) Toluene-d8	8.647	98	302254	48.576	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	97.160%		
62) 4-Bromofluorobenzene	11.079	95	110625	45.240	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	90.480%		
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.325	101	3825	2.306	ug/l	94
12) 1,1-Dichloroethene	2.312	96	757	0.469	ug/l #	76
27) cis-1,2-Dichloroethene	4.489	96	2594	1.198	ug/l	73
44) Trichloroethene	7.123	130	267971	135.207	ug/l	97
64) Tetrachloroethene	9.275	164	6101	3.656	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082823\
Data File : VX037322.D
Acq On : 29 Aug 2023 05:58
Operator : JC/MD
Sample : 04126-21
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20230821

Quant Time: Aug 29 07:21:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

