

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082823\
 Data File : VX037323.D
 Acq On : 29 Aug 2023 06:22
 Operator : JC/MD
 Sample : 04126-22
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D3-20230821

Quant Time: Aug 29 07:22:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	151529	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	249303	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	221636	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	100929	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	93810	42.942	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.880%
35) Dibromofluoromethane	5.385	113	76599	46.393	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.780%
50) Toluene-d8	8.646	98	291842	48.729	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.460%
62) 4-Bromofluorobenzene	11.079	95	107218	45.554	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.100%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.324	101	3126	1.952	ug/l	96
12) 1,1-Dichloroethene	2.318	96	608	0.390	ug/l #	66
27) cis-1,2-Dichloroethene	4.495	96	1215	0.581	ug/l	84
44) Trichloroethene	7.122	130	107056	56.120	ug/l	99
64) Tetrachloroethene	9.274	164	3055	1.889	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082823\
Data File : VX037323.D
Acq On : 29 Aug 2023 06:22
Operator : JC/MD
Sample : 04126-22
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D3-20230821

Quant Time: Aug 29 07:22:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

