



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082823\
Data File : VX037284.D
Acq On : 28 Aug 2023 08:16
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Title : SW846 8260
Last Update : Fri Aug 18 09:30:38 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.4	8693	PASS
75	95	30	60	50.5	26709	PASS
95	95	100	100	100.0	52925	PASS
96	95	5	9	6.9	3630	PASS
173	174	0.00	2	0.8	319	PASS
174	95	50	100	73.0	38659	PASS
175	174	5	9	7.4	2857	PASS
176	174	95	101	96.5	37296	PASS
177	176	5	9	6.7	2486	PASS