

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082825\
 Data File : VX047547.D
 Acq On : 28 Aug 2025 13:46
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Aug 29 02:37:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 29 02:31:18 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	292249	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	584810	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	575786	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	283812	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	225350	50.213	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.420%
35) Dibromofluoromethane	5.397	113	192496	47.959	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.920%
50) Toluene-d8	8.647	98	674753	47.960	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.920%
62) 4-Bromofluorobenzene	11.079	95	276088	52.236	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082825\
Data File : VX047547.D
Acq On : 28 Aug 2025 13:46
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Aug 29 02:37:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082825W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 29 02:31:18 2025
Response via : Initial Calibration

