

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082918\
 Data File : VX004279.D
 Acq On : 29 Aug 2018 09:10
 Operator : JC/MD
 Sample : VX0829MBL01
 Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0829MBL01

Quant Time: Aug 30 03:35:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 23 05:39:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	300516	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	470102	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	427059	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	238458	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	204864	52.30	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.60%	
35) Dibromofluoromethane	5.50	113	177983	50.89	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.78%	
50) Toluene-d8	8.71	98	656656	49.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.48%	
62) 4-Bromofluorobenzene	11.14	95	231862	47.07	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.14%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082918\
Data File : VX004279.D
Acq On : 29 Aug 2018 09:10
Operator : JC/MD
Sample : VX0829MBL01
Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0829MBL01

Quant Time: Aug 30 03:35:26 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 23 05:39:44 2018
Response via : Initial Calibration

